



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651



REF TEK 147A-01/3

S/N: 14911

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0017	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9949	0.9951	0.9953
30.00	0.9874	0.9880	0.9886
40.00	0.9790	0.9800	0.9811
50.00	0.9715	0.9729	0.9746
60.00	0.9660	0.9679	0.9703
70.00	0.9643	0.9667	0.9698
80.00	0.9677	0.9706	0.9748
90.00	0.9769	0.9802	0.9857
100.00	0.9926	0.9963	1.0035
120.00	1.0441	1.0481	1.0613
140.00	1.1023	1.1044	1.1286
160.00	1.0903	1.0862	1.1194

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.499
Cross-axis Coupling (g/g)	0.08%	0.52%	0.56%