



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3

S/N: 14915

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0024	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9945	0.9934	0.9943
30.00	0.9868	0.9839	0.9863
40.00	0.9785	0.9734	0.9774
50.00	0.9716	0.9642	0.9699
60.00	0.9678	0.9580	0.9653
70.00	0.9687	0.9568	0.9654
80.00	0.9764	0.9622	0.9719
90.00	0.9911	0.9750	0.9854
100.00	1.0140	0.9962	1.0068
120.00	1.0849	1.0654	1.0745
140.00	1.1594	1.1451	1.1481
160.00	1.1308	1.1289	1.1247

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.502	2.501
Cross-axis Coupling (g/g)	0.72%	0.32%	0.32%