



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topples Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14919

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0013	1.0015	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9963	0.9958	0.9954
30.00	0.9911	0.9897	0.9889
40.00	0.9854	0.9832	0.9815
50.00	0.9807	0.9778	0.9751
60.00	0.9781	0.9748	0.9707
70.00	0.9790	0.9759	0.9701
80.00	0.9849	0.9828	0.9746
90.00	0.9963	0.9958	0.9849
100.00	1.0142	1.0161	1.0019
120.00	1.0712	1.0806	1.0588
140.00	1.1406	1.1558	1.1330
160.00	1.1458	1.1478	1.1513

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.503	2.500
Cross-axis Coupling (g/g)	0.40%	0.64%	0.24%