



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14925

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0014	1.0016	1.0014
10.00	1.0000	1.0000	1.0000
20.00	0.9961	0.9955	0.9961
30.00	0.9906	0.9891	0.9904
40.00	0.9845	0.9819	0.9839
50.00	0.9794	0.9756	0.9780
60.00	0.9764	0.9714	0.9739
70.00	0.9769	0.9709	0.9728
80.00	0.9825	0.9756	0.9760
90.00	0.9936	0.9861	0.9841
100.00	1.0111	1.0032	0.9979
120.00	1.0663	1.0596	1.0448
140.00	1.1294	1.1289	1.1077
160.00	1.1222	1.1325	1.1297

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.499	2.502
Cross-axis Coupling (g/g)	0.44%	0.48%	0.48%