The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 14929

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0014	1.0017	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9959	0.9952	0.9953
30.00	0.9902	0.9884	0.9884
40.00	0.9840	0.9809	0.9808
50.00	0.9790	0.9745	0.9741
60.00	0.9764	0.9705	0.9695
70.00	0.9777	0.9705	0.9686
80.00	0.9845	0.9760	0.9729
90.00	0.9970	0.9875	0.9828
100.00	1.0162	1.0057	0.9992
120.00	1.0753	1.0627	1.0533
140.00	1.1409	1.1257	1.1209
160.00	1.1325	1.1153	1.1311

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.499	2.501
Cross-axis Coupling (g/g)	0.12%	0.24%	0.16%