



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14931

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0018	1.0013
10.00	1.0000	1.0000	1.0000
20.00	0.9955	0.9949	0.9965
30.00	0.9891	0.9877	0.9915
40.00	0.9821	0.9796	0.9860
50.00	0.9761	0.9726	0.9815
60.00	0.9726	0.9679	0.9790
70.00	0.9731	0.9673	0.9801
80.00	0.9794	0.9723	0.9861
90.00	0.9920	0.9837	0.9976
100.00	1.0120	1.0025	1.0157
120.00	1.0765	1.0643	1.0744
140.00	1.1521	1.1409	1.1517
160.00	1.1412	1.1447	1.1741

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.498	2.499	2.501
Cross-axis Coupling (g/g)	0.28%	0.04%	0.04%