



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14943

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0017	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9954	0.9951	0.9945
30.00	0.9890	0.9881	0.9867
40.00	0.9819	0.9804	0.9779
50.00	0.9759	0.9737	0.9703
60.00	0.9723	0.9694	0.9652
70.00	0.9726	0.9691	0.9644
80.00	0.9784	0.9742	0.9693
90.00	0.9899	0.9853	0.9806
100.00	1.0079	1.0031	0.9990
120.00	1.0633	1.0597	1.0584
140.00	1.1218	1.1250	1.1265
160.00	1.1056	1.1216	1.1187

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.499
Cross-axis Coupling (g/g)	0.24%	0.16%	0.04%