The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 14946

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0021	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9947	0.9941	0.9950
30.00	0.9872	0.9855	0.9878
40.00	0.9789	0.9759	0.9799
50.00	0.9719	0.9673	0.9731
60.00	0.9676	0.9609	0.9687
70.00	0.9679	0.9589	0.9684
80.00	0.9743	0.9625	0.9739
90.00	0.9875	0.9726	0.9856
100.00	1.0085	0.9901	1.0046
120.00	1.0754	1.0500	1.0664
140.00	1.1530	1.1291	1.1427
160.00	1.1433	1.1467	1.1484

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.502
Cross-axis Coupling (g/g)	0.08%	0.04%	0.04%