



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14951

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0023	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9946	0.9935	0.9949
30.00	0.9871	0.9841	0.9877
40.00	0.9789	0.9737	0.9797
50.00	0.9724	0.9645	0.9728
60.00	0.9690	0.9580	0.9685
70.00	0.9706	0.9564	0.9685
80.00	0.9792	0.9613	0.9745
90.00	0.9952	0.9734	0.9870
100.00	1.0199	0.9938	1.0073
120.00	1.0968	1.0621	1.0740
140.00	1.1792	1.1475	1.1607
160.00	1.1465	1.1492	1.1774

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.502	2.502
Cross-axis Coupling (g/g)	0.04%	0.44%	0.04%