



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14952

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0020	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9946	0.9943	0.9939
30.00	0.9869	0.9863	0.9852
40.00	0.9785	0.9775	0.9754
50.00	0.9715	0.9699	0.9666
60.00	0.9673	0.9652	0.9602
70.00	0.9677	0.9652	0.9581
80.00	0.9745	0.9716	0.9616
90.00	0.9880	0.9849	0.9717
100.00	1.0092	1.0064	0.9890
120.00	1.0751	1.0751	1.0477
140.00	1.1447	1.1540	1.1215
160.00	1.1205	1.1399	1.1285

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.500
Cross-axis Coupling (g/g)	0.72%	0.08%	0.12%