



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14955

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0015	1.0017	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9957	0.9952	0.9949
30.00	0.9896	0.9883	0.9876
40.00	0.9828	0.9808	0.9796
50.00	0.9771	0.9746	0.9729
60.00	0.9737	0.9710	0.9689
70.00	0.9744	0.9720	0.9694
80.00	0.9807	0.9792	0.9763
90.00	0.9933	0.9934	0.9900
100.00	1.0135	1.0157	1.0119
120.00	1.0804	1.0884	1.0828
140.00	1.1679	1.1778	1.1678
160.00	1.1809	1.1755	1.1622

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.501
Cross-axis Coupling (g/g)	0.04%	0.04%	0.64%