



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14956

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0015	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9947	0.9958	0.9950
30.00	0.9871	0.9899	0.9879
40.00	0.9789	0.9834	0.9800
50.00	0.9721	0.9781	0.9732
60.00	0.9682	0.9750	0.9688
70.00	0.9691	0.9758	0.9686
80.00	0.9767	0.9819	0.9741
90.00	0.9914	0.9938	0.9860
100.00	1.0144	1.0125	1.0053
120.00	1.0877	1.0721	1.0683
140.00	1.1707	1.1462	1.1462
160.00	1.1504	1.1602	1.1483

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.502
Cross-axis Coupling (g/g)	0.40%	0.00%	0.36%