The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 14958

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0021	1.0018	1.0019
10.00	1.0000	1.0000	1.0000
20.00	0.9942	0.9950	0.9947
30.00	0.9861	0.9879	0.9871
40.00	0.9771	0.9802	0.9788
50.00	0.9695	0.9736	0.9718
60.00	0.9647	0.9695	0.9675
70.00	0.9647	0.9697	0.9678
80.00	0.9712	0.9759	0.9743
90.00	0.9847	0.9885	0.9877
100.00	1.0065	1.0086	1.0091
120.00	1.0759	1.0740	1.0785
140.00	1.1538	1.1555	1.1641
160.00	1.1342	1.1639	1.1682

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.499
Cross-axis Coupling (g/g)	0.16%	0.16%	0.16%