The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 14960

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0018	1.0015
10.00	1.0000	1.0000	1.0000
20.00	0.9955	0.9950	0.9959
30.00	0.9889	0.9876	0.9900
40.00	0.9816	0.9793	0.9833
50.00	0.9751	0.9716	0.9774
60.00	0.9706	0.9658	0.9733
70.00	0.9698	0.9636	0.9726
80.00	0.9740	0.9662	0.9765
90.00	0.9840	0.9744	0.9857
100.00	1.0006	0.9891	1.0011
120.00	1.0556	1.0401	1.0523
140.00	1.1242	1.1078	1.1191
160.00	1.1292	1.1237	1.1355

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.500
Cross-axis Coupling (g/g)	0.28%	0.04%	0.24%