



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14962

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0018	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9951	0.9949	0.9954
30.00	0.9880	0.9874	0.9888
40.00	0.9802	0.9791	0.9816
50.00	0.9735	0.9717	0.9754
60.00	0.9691	0.9665	0.9716
70.00	0.9689	0.9653	0.9719
80.00	0.9746	0.9698	0.9781
90.00	0.9867	0.9806	0.9907
100.00	1.0066	0.9991	1.0109
120.00	1.0722	1.0628	1.0780
140.00	1.1550	1.1492	1.1651
160.00	1.1592	1.1684	1.1778

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.503	2.502	2.503
Cross-axis Coupling (g/g)	0.12%	0.28%	0.64%