



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14963

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0015	1.0014
10.00	1.0000	1.0000	1.0000
20.00	0.9952	0.9957	0.9960
30.00	0.9884	0.9895	0.9902
40.00	0.9809	0.9824	0.9838
50.00	0.9746	0.9762	0.9782
60.00	0.9709	0.9719	0.9745
70.00	0.9715	0.9711	0.9742
80.00	0.9780	0.9751	0.9787
90.00	0.9908	0.9844	0.9885
100.00	1.0109	1.0001	1.0045
120.00	1.0730	1.0522	1.0573
140.00	1.1365	1.1195	1.1244
160.00	1.1065	1.1352	1.1372

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.503	2.500	2.499
Cross-axis Coupling (g/g)	0.04%	0.12%	0.32%