The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 14969

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0016	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9952	0.9956	0.9951
30.00	0.9883	0.9894	0.9883
40.00	0.9809	0.9826	0.9808
50.00	0.9749	0.9769	0.9745
60.00	0.9716	0.9736	0.9709
70.00	0.9728	0.9741	0.9715
80.00	0.9801	0.9800	0.9780
90.00	0.9939	0.9915	0.9908
100.00	1.0152	1.0094	1.0108
120.00	1.0809	1.0644	1.0729
140.00	1.1517	1.1218	1.1415
160.00	1.1337	1.1034	1.1306

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.502
Cross-axis Coupling (g/g)	0.48%	0.32%	0.28%