



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14976

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0018	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9952	0.9949	0.9952
30.00	0.9884	0.9876	0.9883
40.00	0.9808	0.9793	0.9808
50.00	0.9743	0.9720	0.9743
60.00	0.9699	0.9667	0.9702
70.00	0.9695	0.9652	0.9701
80.00	0.9742	0.9688	0.9755
90.00	0.9847	0.9781	0.9868
100.00	1.0016	0.9940	1.0050
120.00	1.0550	1.0474	1.0629
140.00	1.1160	1.1159	1.1319
160.00	1.1121	1.1305	1.1366

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.500	2.502
Cross-axis Coupling (g/g)	0.12%	0.08%	0.20%