



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14978

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0021	1.0015	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9940	0.9956	0.9954
30.00	0.9854	0.9894	0.9888
40.00	0.9758	0.9825	0.9818
50.00	0.9673	0.9767	0.9760
60.00	0.9614	0.9730	0.9729
70.00	0.9598	0.9731	0.9740
80.00	0.9642	0.9782	0.9811
90.00	0.9750	0.9889	0.9944
100.00	0.9932	1.0059	1.0149
120.00	1.0522	1.0590	1.0779
140.00	1.1197	1.1190	1.1469
160.00	1.1105	1.1139	1.1349

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.499
Cross-axis Coupling (g/g)	0.32%	0.12%	0.40%