



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 14980

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0020	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9949	0.9944	0.9940
30.00	0.9877	0.9864	0.9855
40.00	0.9797	0.9774	0.9760
50.00	0.9727	0.9695	0.9677
60.00	0.9681	0.9640	0.9620
70.00	0.9676	0.9628	0.9610
80.00	0.9725	0.9674	0.9660
90.00	0.9832	0.9783	0.9777
100.00	1.0006	0.9964	0.9970
120.00	1.0554	1.0557	1.0592
140.00	1.1153	1.1255	1.1290
160.00	1.1041	1.1203	1.1137

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.502
Cross-axis Coupling (g/g)	0.44%	0.44%	0.52%