



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17704

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0016	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9948	0.9955	0.9950
30.00	0.9874	0.9889	0.9877
40.00	0.9791	0.9816	0.9795
50.00	0.9717	0.9752	0.9723
60.00	0.9667	0.9710	0.9672
70.00	0.9657	0.9704	0.9662
80.00	0.9703	0.9751	0.9706
90.00	0.9811	0.9856	0.9813
100.00	0.9991	1.0028	0.9994
120.00	1.0590	1.0598	1.0610
140.00	1.1339	1.1307	1.1408
160.00	1.1407	1.1370	1.1488

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.501	2.499
Cross-axis Coupling (g/g)	0.08%	0.52%	0.60%