



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17710

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0014	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9948	0.9961	0.9951
30.00	0.9874	0.9906	0.9881
40.00	0.9790	0.9844	0.9803
50.00	0.9715	0.9792	0.9734
60.00	0.9662	0.9761	0.9687
70.00	0.9648	0.9765	0.9680
80.00	0.9688	0.9820	0.9726
90.00	0.9790	0.9928	0.9835
100.00	0.9966	1.0101	1.0014
120.00	1.0568	1.0658	1.0618
140.00	1.1391	1.1357	1.1416
160.00	1.1633	1.1489	1.1615

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.502
Cross-axis Coupling (g/g)	0.16%	0.68%	0.76%