



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17716

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0016	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9954	0.9956	0.9953
30.00	0.9889	0.9894	0.9886
40.00	0.9816	0.9827	0.9811
50.00	0.9752	0.9771	0.9748
60.00	0.9709	0.9739	0.9709
70.00	0.9703	0.9749	0.9711
80.00	0.9748	0.9816	0.9768
90.00	0.9852	0.9945	0.9887
100.00	1.0023	1.0148	1.0079
120.00	1.0600	1.0805	1.0705
140.00	1.1413	1.1654	1.1528
160.00	1.1815	1.1849	1.1781

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.502
Cross-axis Coupling (g/g)	0.08%	0.60%	0.44%