The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 17718

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0014	1.0015
10.00	1.0000	1.0000	1.0000
20.00	0.9955	0.9961	0.9959
30.00	0.9892	0.9907	0.9901
40.00	0.9824	0.9849	0.9837
50.00	0.9769	0.9803	0.9783
60.00	0.9740	0.9781	0.9752
70.00	0.9753	0.9797	0.9758
80.00	0.9824	0.9868	0.9815
90.00	0.9956	0.9994	0.9927
100.00	1.0159	1.0184	1.0104
120.00	1.0779	1.0752	1.0659
140.00	1.1445	1.1326	1.1323
160.00	1.1303	1.1109	1.1391

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.500	2.499
Cross-axis Coupling (g/g)	0.24%	0.28%	0.40%