



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17720

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0024	1.0020	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9931	0.9944	0.9941
30.00	0.9833	0.9865	0.9857
40.00	0.9723	0.9777	0.9764
50.00	0.9625	0.9703	0.9681
60.00	0.9554	0.9655	0.9625
70.00	0.9531	0.9653	0.9614
80.00	0.9571	0.9713	0.9663
90.00	0.9683	0.9840	0.9779
100.00	0.9874	1.0044	0.9973
120.00	1.0513	1.0690	1.0617
140.00	1.1267	1.1408	1.1414
160.00	1.1157	1.1249	1.1444

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.500
Cross-axis Coupling (g/g)	0.24%	0.28%	0.68%