



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17732

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0021	1.0019
10.00	1.0000	1.0000	1.0000
20.00	0.9950	0.9942	0.9948
30.00	0.9880	0.9859	0.9874
40.00	0.9803	0.9766	0.9793
50.00	0.9740	0.9685	0.9726
60.00	0.9706	0.9631	0.9687
70.00	0.9718	0.9623	0.9697
80.00	0.9798	0.9679	0.9775
90.00	0.9950	0.9806	0.9926
100.00	1.0189	1.0017	1.0166
120.00	1.0960	1.0716	1.0940
140.00	1.1863	1.1544	1.1821
160.00	1.1663	1.1364	1.1515

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.501
Cross-axis Coupling (g/g)	0.24%	0.64%	0.04%