



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17733

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0020	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9944	0.9944	0.9943
30.00	0.9863	0.9863	0.9860
40.00	0.9774	0.9773	0.9768
50.00	0.9695	0.9693	0.9686
60.00	0.9643	0.9638	0.9628
70.00	0.9635	0.9627	0.9615
80.00	0.9690	0.9677	0.9662
90.00	0.9815	0.9795	0.9778
100.00	1.0022	0.9995	0.9975
120.00	1.0717	1.0678	1.0644
140.00	1.1589	1.1583	1.1497
160.00	1.1558	1.1707	1.1511

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.501	2.501
Cross-axis Coupling (g/g)	0.08%	0.04%	0.32%