The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 17739

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0021	1.0024	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9942	0.9933	0.9940
30.00	0.9859	0.9838	0.9855
40.00	0.9767	0.9732	0.9759
50.00	0.9688	0.9638	0.9675
60.00	0.9635	0.9572	0.9616
70.00	0.9629	0.9556	0.9602
80.00	0.9687	0.9607	0.9650
90.00	0.9818	0.9733	0.9767
100.00	1.0035	0.9947	0.9964
120.00	1.0761	1.0669	1.0626
140.00	1.1679	1.1555	1.1443
160.00	1.1670	1.1456	1.1422

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.502
Cross-axis Coupling (g/g)	0.08%	0.28%	0.56%