



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17741

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0017	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9948	0.9951	0.9944
30.00	0.9874	0.9881	0.9863
40.00	0.9791	0.9803	0.9772
50.00	0.9718	0.9738	0.9691
60.00	0.9670	0.9698	0.9635
70.00	0.9664	0.9701	0.9624
80.00	0.9718	0.9766	0.9672
90.00	0.9839	0.9899	0.9790
100.00	1.0041	1.0114	0.9990
120.00	1.0720	1.0823	1.0669
140.00	1.1589	1.1714	1.1544
160.00	1.1603	1.1701	1.1557

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.501
Cross-axis Coupling (g/g)	0.08%	0.04%	0.64%