



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17750

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0021	1.0022
10.00	1.0000	1.0000	1.0000
20.00	0.9944	0.9941	0.9939
30.00	0.9865	0.9857	0.9852
40.00	0.9780	0.9763	0.9755
50.00	0.9708	0.9681	0.9672
60.00	0.9666	0.9625	0.9618
70.00	0.9673	0.9614	0.9611
80.00	0.9748	0.9666	0.9670
90.00	0.9896	0.9787	0.9802
100.00	1.0132	0.9989	1.0017
120.00	1.0901	1.0671	1.0726
140.00	1.1853	1.1549	1.1593
160.00	1.1849	1.1621	1.1574

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.501	2.501
Cross-axis Coupling (g/g)	0.04%	0.28%	0.08%