



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17751

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0023	1.0021	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9935	0.9940	0.9942
30.00	0.9842	0.9855	0.9860
40.00	0.9738	0.9764	0.9770
50.00	0.9647	0.9688	0.9693
60.00	0.9585	0.9644	0.9644
70.00	0.9571	0.9653	0.9643
80.00	0.9623	0.9733	0.9707
90.00	0.9750	0.9889	0.9843
100.00	0.9962	1.0135	1.0063
120.00	1.0677	1.0914	1.0782
140.00	1.1571	1.1768	1.1659
160.00	1.1543	1.1473	1.1628

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.501
Cross-axis Coupling (g/g)	0.08%	0.48%	0.16%