The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 17758

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0021	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9947	0.9942	0.9944
30.00	0.9873	0.9860	0.9865
40.00	0.9792	0.9769	0.9778
50.00	0.9724	0.9691	0.9703
60.00	0.9684	0.9640	0.9654
70.00	0.9691	0.9637	0.9652
80.00	0.9762	0.9698	0.9712
90.00	0.9904	0.9830	0.9842
100.00	1.0128	1.0044	1.0052
120.00	1.0854	1.0749	1.0741
140.00	1.1738	1.1624	1.1593
160.00	1.1716	1.1645	1.1605

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.499
Cross-axis Coupling (g/g)	0.36%	0.24%	0.08%