



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17769

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0025	1.0025	1.0025
10.00	1.0000	1.0000	1.0000
20.00	0.9931	0.9929	0.9931
30.00	0.9831	0.9827	0.9832
40.00	0.9719	0.9711	0.9723
50.00	0.9617	0.9607	0.9627
60.00	0.9541	0.9528	0.9562
70.00	0.9513	0.9497	0.9548
80.00	0.9547	0.9527	0.9603
90.00	0.9651	0.9627	0.9735
100.00	0.9836	0.9808	0.9955
120.00	1.0474	1.0432	1.0691
140.00	1.1288	1.1228	1.1595
160.00	1.1330	1.1281	1.1544

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.501
Cross-axis Coupling (g/g)	0.20%	0.60%	0.52%