



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17773

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0032	1.0036	1.0034
10.00	1.0000	1.0000	1.0000
20.00	0.9913	0.9903	0.9907
30.00	0.9798	0.9774	0.9780
40.00	0.9686	0.9643	0.9649
50.00	0.9605	0.9542	0.9544
60.00	0.9571	0.9487	0.9482
70.00	0.9600	0.9495	0.9480
80.00	0.9704	0.9578	0.9553
90.00	0.9880	0.9733	0.9699
100.00	1.0135	0.9967	0.9926
120.00	1.0885	1.0668	1.0636
140.00	1.1696	1.1434	1.1497
160.00	1.1533	1.1286	1.1549

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.500
Cross-axis Coupling (g/g)	0.32%	0.44%	0.76%