The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 17774

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0036	1.0046	1.0037
10.00	1.0000	1.0000	1.0000
20.00	0.9902	0.9882	0.9903
30.00	0.9773	0.9739	0.9779
40.00	0.9644	0.9607	0.9661
50.00	0.9547	0.9512	0.9576
60.00	0.9497	0.9463	0.9538
70.00	0.9509	0.9470	0.9558
80.00	0.9597	0.9545	0.9650
90.00	0.9757	0.9686	0.9812
100.00	0.9996	0.9903	1.0053
120.00	1.0715	1.0582	1.0789
140.00	1.1501	1.1348	1.1637
160.00	1.1330	1.1131	1.1440

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.499	2.502
Cross-axis Coupling (g/g)	0.44%	0.48%	0.08%