The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 17775

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0019	1.0022
10.00	1.0000	1.0000	1.0000
20.00	0.9939	0.9947	0.9939
30.00	0.9853	0.9872	0.9852
40.00	0.9757	0.9790	0.9757
50.00	0.9675	0.9720	0.9676
60.00	0.9623	0.9677	0.9624
70.00	0.9619	0.9678	0.9623
80.00	0.9682	0.9741	0.9689
90.00	0.9820	0.9871	0.9831
100.00	1.0044	1.0080	1.0062
120.00	1.0791	1.0756	1.0821
140.00	1.1755	1.1578	1.1761
160.00	1.1864	1.1573	1.1734

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.502
Cross-axis Coupling (g/g)	0.04%	0.32%	0.56%