



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 17777

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0024	1.0022	1.0024
10.00	1.0000	1.0000	1.0000
20.00	0.9934	0.9937	0.9932
30.00	0.9839	0.9848	0.9834
40.00	0.9733	0.9751	0.9725
50.00	0.9638	0.9667	0.9627
60.00	0.9570	0.9614	0.9556
70.00	0.9549	0.9611	0.9534
80.00	0.9589	0.9678	0.9575
90.00	0.9700	0.9819	0.9688
100.00	0.9889	1.0047	0.9884
120.00	1.0525	1.0783	1.0548
140.00	1.1310	1.1609	1.1380
160.00	1.1333	1.1349	1.1393

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.500
Cross-axis Coupling (g/g)	0.44%	0.44%	0.44%