The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 17779

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0026	1.0026	1.0025
10.00	1.0000	1.0000	1.0000
20.00	0.9928	0.9928	0.9930
30.00	0.9824	0.9824	0.9831
40.00	0.9708	0.9707	0.9721
50.00	0.9602	0.9603	0.9624
60.00	0.9523	0.9528	0.9558
70.00	0.9493	0.9502	0.9544
80.00	0.9526	0.9541	0.9599
90.00	0.9631	0.9654	0.9732
100.00	0.9819	0.9852	0.9955
120.00	1.0467	1.0523	1.0698
140.00	1.1281	1.1355	1.1592
160.00	1.1258	1.1342	1.1463

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.501
Cross-axis Coupling (g/g)	0.24%	0.24%	0.40%