The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 18702

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9989	0.9995	0.9990
10.00	1.0000	1.0000	1.0000
20.00	1.0021	1.0012	1.0025
30.00	1.0035	1.0024	1.0051
40.00	1.0034	1.0027	1.0069
50.00	1.0021	1.0019	1.0073
60.00	1.0002	1.0003	1.0064
70.00	0.9986	0.9985	1.0049
80.00	0.9978	0.9972	1.0035
90.00	0.9985	0.9974	1.0032
100.00	1.0015	0.9997	1.0046
120.00	1.0168	1.0143	1.0160
140.00	1.0494	1.0470	1.0422
160.00	1.0940	1.0904	1.0762

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.500	2.501
Cross-axis Coupling (g/g)	0.20%	0.08%	0.52%