



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 18706

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0001	1.0000	1.0003
10.00	1.0000	1.0000	1.0000
20.00	0.9996	1.0001	0.9991
30.00	0.9989	1.0003	0.9978
40.00	0.9978	1.0005	0.9961
50.00	0.9965	1.0008	0.9942
60.00	0.9948	1.0013	0.9923
70.00	0.9931	1.0021	0.9907
80.00	0.9915	1.0036	0.9898
90.00	0.9904	1.0058	0.9902
100.00	0.9904	1.0095	0.9922
120.00	0.9960	1.0234	1.0042
140.00	1.0159	1.0514	1.0316
160.00	1.0515	1.0903	1.0697

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.499
Cross-axis Coupling (g/g)	0.96%	0.16%	0.68%