The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 18716

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0033	1.0033	1.0031
10.00	1.0000	1.0000	1.0000
20.00	0.9910	0.9910	0.9915
30.00	0.9791	0.9789	0.9798
40.00	0.9670	0.9667	0.9677
50.00	0.9578	0.9573	0.9578
60.00	0.9532	0.9524	0.9519
70.00	0.9547	0.9535	0.9517
80.00	0.9638	0.9619	0.9584
90.00	0.9800	0.9773	0.9719
100.00	1.0042	1.0003	0.9929
120.00	1.0767	1.0693	1.0573
140.00	1.1569	1.1459	1.1319
160.00	1.1451	1.1363	1.1334

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.499
Cross-axis Coupling (g/g)	0.24%	0.40%	0.52%