



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 18718

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0030	1.0029	1.0030
10.00	1.0000	1.0000	1.0000
20.00	0.9917	0.9921	0.9918
30.00	0.9805	0.9814	0.9810
40.00	0.9691	0.9704	0.9701
50.00	0.9602	0.9619	0.9619
60.00	0.9556	0.9575	0.9579
70.00	0.9570	0.9588	0.9594
80.00	0.9658	0.9670	0.9675
90.00	0.9818	0.9820	0.9818
100.00	1.0057	1.0043	1.0029
120.00	1.0770	1.0704	1.0649
140.00	1.1550	1.1424	1.1332
160.00	1.1443	1.1369	1.1301

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.500
Cross-axis Coupling (g/g)	0.84%	0.64%	0.04%