



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3

S/N: 18720

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0031	1.0033	1.0033
10.00	1.0000	1.0000	1.0000
20.00	0.9915	0.9910	0.9909
30.00	0.9800	0.9788	0.9788
40.00	0.9681	0.9663	0.9664
50.00	0.9589	0.9565	0.9568
60.00	0.9539	0.9509	0.9516
70.00	0.9550	0.9514	0.9523
80.00	0.9636	0.9590	0.9604
90.00	0.9794	0.9735	0.9755
100.00	1.0031	0.9956	0.9984
120.00	1.0736	1.0622	1.0683
140.00	1.1482	1.1362	1.1495
160.00	1.1313	1.1283	1.1481

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.499
Cross-axis Coupling (g/g)	0.16%	0.48%	0.12%