The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 18728

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0025	1.0023	1.0023
10.00	1.0000	1.0000	1.0000
20.00	0.9929	0.9935	0.9936
30.00	0.9829	0.9842	0.9846
40.00	0.9721	0.9740	0.9748
50.00	0.9631	0.9652	0.9665
60.00	0.9575	0.9593	0.9612
70.00	0.9574	0.9584	0.9611
80.00	0.9647	0.9639	0.9679
90.00	0.9798	0.9762	0.9821
100.00	1.0037	0.9963	1.0050
120.00	1.0779	1.0605	1.0788
140.00	1.1557	1.1362	1.1649
160.00	1.1247	1.1379	1.1541

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.501
Cross-axis Coupling (g/g)	0.08%	0.04%	0.84%