The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3
S/N: 18741

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0021	1.0023
10.00	1.0000	1.0000	1.0000
20.00	0.9939	0.9940	0.9936
30.00	0.9851	0.9854	0.9845
40.00	0.9754	0.9757	0.9743
50.00	0.9668	0.9671	0.9651
60.00	0.9608	0.9610	0.9586
70.00	0.9596	0.9593	0.9568
80.00	0.9647	0.9637	0.9612
90.00	0.9769	0.9749	0.9726
100.00	0.9974	0.9941	0.9921
120.00	1.0658	1.0580	1.0570
140.00	1.1475	1.1335	1.1322
160.00	1.1347	1.1189	1.1141

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.500
Cross-axis Coupling (g/g)	0.28%	0.56%	0.48%