The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/03/27

REF TEK 147A-01/3 Accelerometer

S/N:19704 P/N: 97499-00

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0003	1.0002
10.00	1.0000	1.0000	1.0000
20.00	0.9994	0.9993	0.9996
30.00	0.9985	0.9981	0.9989
40.00	0.9973	0.9967	0.9980
50.00	0.9961	0.9953	0.9972
60.00	0.9949	0.9939	0.9965
70.00	0.9941	0.9930	0.9962
80.00	0.9937	0.9927	0.9964
90.00	0.9942	0.9935	0.9976
100.00	0.9957	0.9955	0.9999
120.00	1.0035	1.0045	1.0095
140.00	1.0197	1.0220	1.0280
160.00	1.0410	1.0422	1.0505

SENSITIVITIES

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.500
Cross-axis Coupling (g/g)	0.24%	0.08%	0.40%

* g = 9.80 m/s^2

Rev C