The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/03/27

REF TEK 147A-01/3 Accelerometer

S/N:19705 P/N: 97499-00

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0001	1.0001	1.0002
10.00	1.0000	1.0000	1.0000
20.00	0.9996	0.9998	0.9994
30.00	0.9990	0.9994	0.9984
40.00	0.9981	0.9989	0.9971
50.00	0.9973	0.9984	0.9957
60.00	0.9965	0.9979	0.9944
70.00	0.9960	0.9976	0.9934
80.00	0.9961	0.9978	0.9928
90.00	0.9969	0.9985	0.9932
100.00	0.9988	1.0001	0.9947
120.00	1.0076	1.0075	1.0029
140.00	1.0259	1.0230	1.0210
160.00	1.0507	1.0439	1.0461

SENSITIVITIES

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.500
Cross-axis Coupling (g/g)	0.12%	0.16%	0.08%

* g = 9.80 m/s^2

Rev C