The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/03/28

REF TEK 147A-01/3 Accelerometer

S/N:19708 P/N: 97499-00

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0001	1.0002
10.00	1.0000	1.0000	1.0000
20.00	0.9994	0.9997	0.9994
30.00	0.9985	0.9992	0.9986
40.00	0.9974	0.9986	0.9975
50.00	0.9962	0.9980	0.9964
60.00	0.9950	0.9976	0.9954
70.00	0.9942	0.9974	0.9948
80.00	0.9938	0.9977	0.9949
90.00	0.9942	0.9988	0.9958
100.00	0.9957	1.0008	0.9979
120.00	1.0032	1.0091	1.0072
140.00	1.0186	1.0247	1.0249
160.00	1.0375	1.0432	1.0461

SENSITIVITIES

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.499
Cross-axis Coupling (g/g)	0.28%	0.20%	0.48%

* g = 9.80 m/s^2

Rev C