The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/03/27

REF TEK 147A-01/3 Accelerometer

S/N:19709 P/N: 97499-00

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0000	1.0001	1.0002
10.00	1.0000	1.0000	1.0000
20.00	1.0001	0.9998	0.9994
30.00	1.0002	0.9995	0.9985
40.00	1.0004	0.9991	0.9974
50.00	1.0007	0.9986	0.9962
60.00	1.0012	0.9982	0.9952
70.00	1.0019	0.9979	0.9947
80.00	1.0031	0.9980	0.9948
90.00	1.0048	0.9987	0.9958
100.00	1.0074	1.0002	0.9981
120.00	1.0163	1.0072	1.0076
140.00	1.0321	1.0220	1.0254
160.00	1.0506	1.0419	1.0465

SENSITIVITIES

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.501	2.499
Cross-axis Coupling (g/g)	0.32%	0.32%	0.08%

* g = 9.80 m/s^2

Rev C