The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15101

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0027	1.0021	1.0026
10.00	1.0000	1.0000	1.0000
20.00	0.9924	0.9942	0.9928
30.00	0.9815	0.9859	0.9826
40.00	0.9696	0.9766	0.9713
50.00	0.9592	0.9684	0.9616
60.00	0.9519	0.9629	0.9549
70.00	0.9501	0.9619	0.9536
80.00	0.9552	0.9671	0.9593
90.00	0.9677	0.9792	0.9725
100.00	0.9885	0.9993	0.9943
120.00	1.0545	1.0660	1.0630
140.00	1.1258	1.1462	1.1329
160.00	1.1116	1.1385	1.0989

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.501
Cross-axis Coupling (g/g)	0.08%	0.48%	0.48%